

Certificate

Certificate no.

CU 72503787 0001

License Holder:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Manufacturing Plant:

Analytik Jena GmbH+Co. KG
Konrad-Zuse-Str. 1
07745 Jena
Deutschland

Report Number:

DE25GBQL 001

Client Reference:

Certification acc. to:

UL 61010-1:2012 R6.23
UL 61010-2-061:2019
CSA C22.2 No. 61010-1-12 (R2022) R6.23
CSA C22.2 No. 61010-2-061:19

Product Information

Certified Product:

ICP Emission Spectrometer (ICP-OES)

Model Designation:

PlasmaQuant 9200
PlasmaQuant 9200 Elite

The models differ in terms of software and calculation methods.

Technical Data:

AC 200-240V; 50/60Hz; 2500VA
Protection class I

Remarks:

Trademark: analytikjena

Appendix:

1, 1-12

Date of issue:

2025-09-05
(yr/mo/day)



TUV Rheinland of North America, Inc.
400 Beaver Brook Rd, Boxborough, MA 01719
Tel +1 (978) 266 9500, Fax +1 (978) 266-9992

www.tuv.com

 TÜVRheinland®